

ACDC-400 Instruction Manual Supplement

Title: ACDC-400 Instruction Manual_Rev002

Part Number: 2730810

Print Date: March, 2018

Revision/Date: 20-Mar-2017

Supplement Issue: 1

Issue Date: 3/2018

Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change # 1

On page # 12,
under "SPECIFICATIONS – Continuity Tester" ,
change

From:

Audible threshold: < 40Ω

To:

Audible threshold: ≤ 20Ω beeper ON; > 40Ω beeper OFF